

ELECTROCHEMICAL MAGNETO IMMUNOSENSING

A NOVEL STRATEGY FOR THE DETECTION OF PESTICIDES RESIDUES

E. Zacco, M. I. Pividori, and S. Alegret*

Sensors & Biosensors Group, Analytical Chemistry Division, Department of Chemistry, Autonomous
University of Barcelona, 08193 Bellaterra, Catalonia, Spain

*Department of Analytical Chemistry, University of Barcelona, 08028 Barcelona, Catalonia, Spain.

R. Galve, M.-P. Marco

Applied Molecular Receptors Group (AMRg), IIQAB-CSIC, 08034, Barcelona, Catalonia, Spain.

*Permanent address: Facultad de Bioquímica y Ciencias Biológicas, Universidad Nacional del Litoral,
Santa Fe, Argentina.

*E-mail address: Isabel.Pividori@uab.es

Keywords: Atrazine, Electrochemical Immunosensing, Magnetic Beads, Magneto-ELISA

ABSTRACT

A novel electrochemical immunosensing strategy for the detection of atrazine based on magnetic beads is presented. Different coupling strategies for the modification of the magnetic beads with the specific anti-atrazine antibody have been developed. The immunological reaction for the detection of atrazine performed on the magnetic bead is based on a direct competitive assay using a peroxidase (HRP) tracer as the enzymatic label. After the immunochemical reactions, the modified magnetic beads can be easily captured by a magneto sensor made of graphite-epoxy composite (m-GEC) which is also used as the transducer for the electrochemical immunosensing. The electrochemical detection is thus achieved through a suitable substrate and mediator for the enzyme HRP. The electrochemical approach is also compared with a novel magneto ELISA based on optical detection. The performance of the electrochemical immunosensing strategy based on magnetic beads was successfully evaluated using spiked real orange juice samples. The detection limit for atrazine using the competitive electrochemical magneto immunosensing strategy with anti-atrazine specific antibody covalent coupled with tosyl-activated magnetic beads was found to be $6 \times 10^{-3} \mu\text{g L}^{-1}$ ($0.027 \text{ nmol L}^{-1}$). This strategy offers great promise for rapid, simple, cost-effective and *on-site* analysis of biological, food and environmental samples.